



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

June 13, 2012

MINNTECH CORPORATION
CANTEL MEDICAL COMPANY
TOM ARCAND
14605 28TH AVENUE NORTH
MINNEAPOLIS MN 55447

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: MINNTECH CORPORATION
Product Name: AUTOMATED ENDOSCOPE REPROCESSOR
Model Number(s): MEDIVATORE CER OPTIMA
CER-1 OPTIMA, AND CER-2 OPTIMA
Product File No: 20120224

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2017.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.
- The discharge from this product must discharge to the sanitary drain system through air-gap or air break into a trapped and vented receptor.
- These devices must be installed, maintained and serviced as directed by the manufacturer to perform as advertised.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov